

Bio V[®]

INDICATIONS

Bio V[®] BP Bioactive Glass with DBM Putty is indicated to be gently placed into bony voids or gaps of the skeletal system that are not intrinsic to the stability of the bony structure (i.e., the extremities, pelvis and posterolateral spine). These defects may be surgically created osseous defects or osseous defects created from traumatic injury to the bone. Bio V[®] BP Bioactive Glass with DBM Putty must be used with autograft as a bone graft extender in the posterolateral spine. The product provides a bone graft substitute that remodels into the recipient's skeletal system.

CONTRAINDICATIONS

Bio V[®] BP Bioactive Glass with DBM Putty is contraindicated where the device is intended as structural support in load-bearing bone and in articulating surfaces. Infection at the surgical site and/or distant foci of infections that may spread to the surgical site is a contraindication.

Do not use Bio V[®] BP Bioactive Glass with DBM Putty if package integrity has been compromised.