

DuaFit®

INDICATIONS

The DuaFit® implant is intended for fixation of proximal interphalangeal joint arthro-desis of the lesser toes. Examples include:

- Rigid or semi-rigid hammertoe deformity
- Revision of failed arthroplasty or arthrodesis
- 2nd toe shortening.

CONTRAINDICATIONS

The implant should not be used in a patient who has currently, or who has history of:

- Acute or chronic inflammations, whether local or systemic
- Active infections
- Sensitivity/allergies to the implant materials