

## **RTS® Lesser**

### **INDICATIONS**

The RTS® Lesser MTP Implant is intended for use in the treatment of:

- Partial or complete dislocation of the lesser metatarsophalangeal joint
- Pain associated with either rheumatoid or osteoarthritis
- Repair of unsuccessful arthroplasties of the lesser metatarsophalangeal joint
- Stiffness at the lesser metatarsophalangeal joint associated with joint disease
- Hammertoe deformity where the proximal phalanx is dorsally located on the metatarsal in a fixed contracture state.

### **CONTRAINDICATIONS**

General contraindications for the use of these implants for joint reconstruction, osteotomy or fusion include:

- Significant bone demineralization.
- Inadequate neurovascular status.
- Inadequate skin or musculotendinous system.
- Inadequate bone stock.
- Physiologically or Psychologically unsuitable patient.
- Possibility for conservative treatment.
- Bone, musculature, tendons, or adjacent soft tissue compromised by disease, infection, or prior implantation, which cannot provide adequate support or fixation for the prosthesis.
- Known metal allergy. (1st MPJ w/GrommetsSystem Only)
- Diabetes.
- Active infection.
- Growing patients with open epiphyses.
- Patients with high level of activity