

## **TriBio™**

### **INDICATIONS**

TriBio™ Implant is indicated for use in bony voids or gaps of the skeletal system (i.e. extremities, pelvis and posterolateral spine). These osseous defects are surgically created or result from traumatic injury to the bone and are not intrinsic to the stability of the bony structure. The graft resorbs and is replaced by the growth of new bone during the healing process. TriBio™ Implant must be used with autogenous bone marrow aspirate and autograft in the posterolateral spine.

### **CONTRAINDICATIONS**

TriBio™ Implant is not sold for any use except as indicated. Do not use TriBio™ Implant in the presence of any contraindication.

TriBio™ Implant is contraindicated in patients with a history of severe allergies manifested by a history of anaphylaxis and known allergies to bovine collagen, in patients known to be undergoing desensitization injections to meat products, as these injections can contain bovine collagen, in children and pregnant women, for fractures of the epiphyseal plate, in sites with severe vascular or neurological impairment proximal to the graft site.